

Ref. No.

FR\_RE-1071/0221

BASED ON TEST REPORT ABOUT : VIRUCIDAL EFFECT OF UVC STERILIZERS

## TEST CERTIFICATE

Product

**Sterilizer** (Individual protection and instruments sterilization cabinet)

Model / Type Ref.

VBX25 and VBX50

Ratings and principal characteristics

VBX25 and VBX50 220-240 V, 50-60 Hz,  
VBX25 : 1 x 25W UVC "ViroKill Lamp" Fluocompact (Class I)  
VBX50 : 2 x 25W UVC "ViroKill Lamp" Fluocompact (Class I)

Name and address of the applicant / manufacturer

Test report number

RE-1071/0221

Name and address of the Laboratory doing the test

**Specialised and COFRAC certified virobiology Laboratory**

**FRANCE**

Standard used for the virucidal efficiency test

NF T 72-281

Test performed during period

Mars - Avril 2021

Number of samples of the product tested

1 sample

Number of tests performed

2 tests

Duration of tests

Each test performed during **00h 03min.**

Organisms tested

- *Adenovirus Type 5 (ATCC VR-5)*

- *Norovirus murin S99 (S99 Friedrich Loeffler Institute)*

As shown in the Test Report cited above,  
it has been demonstrated that efficiency  
of the cited devices is at least about :

Organism	Virucidal reduction	Virucidal efficiency
<b>Adenovirus Type 5</b>	5,06 Log reduction	<b>99,99999%</b>
<b>Norovirus murin S99</b>	4,32 Log reduction	<b>99,9999%</b>

Date : 14 / 04 / 2021

Signature :

Company / Fonction :

Name :

